iX10

Patient Monitor

Version 1.0

Main Unit Specification

Physical Specifications

Dimension	260 mm (W) × 227 mm (H) × 155 mm (D)
Weight	< 3.2 kg (standard configuration, excluding battery, accessories, and recorder)

Power Supply

AC Voltage	100 V to 240 V~
Frequency	50 Hz/60 Hz
Input Current	1.6 A to 0.8 A
Over Current Fuse Protection	Support
DC Voltage	12/24V
Input Current	< 3.5 A

Battery

Battery Type	Rechargeable lithium-ion battery	
Operating Time	One battery (2500 mAh)	≥ 3.5 h
	One battery (5000 mAh)	≥ 7 h
Charge Time	One battery (2500 mAh)	≤ 2.5 h (monitor is off) ≤ 5 h (monitor is running or in standby mode)
	One battery (5000 mAh)	≤ 5 h (monitor is off) ≤ 10 h (monitor is running or in standby mode)

Display

Display screen	10.1 inch color TFT, supporting touch screen
Resolution	1280 × 800
Wave	A maximum of 8 waveforms

Recorder

Record Width	48 mm
Paper Speed	12.5 mm/s, 25 mm/s, 50 mm/s
Waveform Channels	Maximum 3 channels
Recording types	Continual real-time recording 8-second real-time recording Trend graph recording Trend table recording C.O. measurement recording NIBP trigger recording ST VIEW recording QT VIEW recording

Data Storage

Trend data	2400 hours @ 1 second
NIBP Measurement	1200 sets
Alarm Events	1000 sets

Wi-Fi

IEEE	802.11a/b/g/n
Frequency Band	2.4 GHz ISM band & 5 G ISM band

Interfaces and others

Nurse Call	1
USB Interfaces	4
HDMI Interface	1
RS232 Interface	1
Wired Network Interface	1

ECG

Lead Mode	3 Electrodes: I, II, III 5 Electrodes: I, II, III, aVR, aVL, aVF, V 6 Electrodes: I, II, III, aVR, aVL, aVF, Va, Vb 10 Electrodes: I, II, III, aVR, aVL, aVF, V1-V6
Electrode Standard	AHA, IEC
Display Sensitivity	×0.125, ×0.25, ×0.5, ×1, ×2, ×4, AUTO
Sweep	6.25 mm/s, 12.5 mm/s, 25 mm/s, 50 mm/s
Bandwidth (-3 dB)	Diagnosis: 0.05 Hz to 150 Hz Diagnosis 1: 0.05 Hz to 40 Hz Monitor: 0.5 Hz to 40 Hz Surgery: 1 Hz to 20 Hz Enhanced: 2 Hz ~18 Hz Customized: High-pass Filter and Low-pass Filter
CMRR	Diagnosis: > 95 dB Diagnosis 1: > 105 dB (when Notch is turned on) Monitor: > 105 dB Surgery: > 105 dB Enhanced: > 105 dB Customized: > 105 dB (Low-pass Filter < 40 Hz) > 95 dB (Low-pass Filter > 40 Hz)
Hum Filter	In diagnosis, Diagnosis 1, monitor, surgery, enhanced and customized modes: 50 Hz/60 Hz (Hum Filter can be turned on or off manually)
Recovery Time After Defibrillation	< 5 s
ESU Protection	Cut mode: 300 W Coagulation mode: 100 W Restore time: ≤10 s
Pace Pulse Detecting Lead	One among I, II, III, aVR, aVL, aVF, V1-V6

EDAN

Heart Rate

Range	ADU: 15 bpm to 300 bpm PED/NEO: 15 bpm to 350 bpm
Accuracy	±1% or ±1 bpm, whichever is greater
Resolution	1 bpm

PVC

Range	ADU: (0 to 300) PVCs/ min PED/NEO: (0 to 350) PVCs/ min
Resolution	1 PVCs/min

Pause/min

Range	ADU/PED/NEO: (0 to 30) pauses/min
Resolution	1 pause/min

ST value

Range	-2.0 mV to +2.0 mV
Accuracy	-0.8 mV to +0.8 mV: ±0.02 mV or 10%, whichever is greater. Beyond this range: not specified.
Resolution	0.01 mV

QT measurement

Range	200 ms ~ 800 ms
Resolution	4 ms
Accuracy	± 30 ms

QTc measurement

Range	200 ms ~ 800 ms
Resolution	1 ms

ΔQTc measurement

Range	-600 ms ~ 600 ms
Resolution	1 ms

Arrhythmia analysis

Asystole, Sustain VT, V-Fib/V-Tach, ExtremeTachy, ExtremeBrady, V-Tach, Vent Brady, Tachy, Brady, Wide QRS Tachy, Non-Sustain VT, Afib, Vent Rhythm, Acc. Vent Rhythm, Pause, Pauses/min High, PVCs High, R on T, PVC Bigeminy, PVC Trigeminy, Pacer not Pacing, Pacer not Capture, Missed Beat, VEB, PVC, Couplet, Run PVCs, IPVC, Irr Rhythm, PAC Bigeminy, Multiform PVCs, PAC Trigeminy, Low Voltage (Limb)

12-lead ECG Synchronization Analysis

Average parameters of heart beat	PR interval (ms)
Heart rate (bpm)	QRS interval (ms)
Time limit of P wave (ms)	QT/QTC (ms)
P-QRS-T AXIS	

RESP

Method	Impedance between RA-LL, RA-LA
Measurement lead	Options are lead I and II. The default is lead II.
RR Measuring Range	0 rpm to 200 rpm
Resolution	1 rpm
Accuracy	6 rpm to 200 rpm: ±2 rpm 0 rpm to 5 rpm: not specified
Gain selection	×0.25, ×0.5, ×1, ×2, ×3, ×4, ×5
Sweep	6.25 mm/s, 12.5 mm/s, 25.0 mm/s, 50.0 mm/s
Apnea Alarm Time	10 s, 15 s, 20 s (Default), 25 s, 30 s, 35 s, 40 s

NIBP

Method	Oscillometry
Mode	Manual, Auto, Continuous, Sequence
Measuring Interval in Auto Mode	1/2/2.5/3/4/5/10/15/30/60/90/120/180/240/360/480 min and User Define
Continuous	5 min, interval is 5 s
Measuring Parameter	SYS, DIA, MAP, PR
Measuring Range	
Adult Mode	SYS: 25 mmHg to 290 mmHg DIA: 10 mmHg to 250 mmHg MAP: 15 mmHg to 260 mmHg
Pediatric Mode	SYS: 25 mmHg to 240 mmHg DIA: 10 mmHg to 200 mmHg MAP: 15 mmHg to 215 mmHg
Neonatal Mode	SYS: 25 mmHg to 140 mmHg DIA: 10 mmHg to 115 mmHg MAP: 15 mmHg to 125 mmHg
Cuff Pressure Measuring Range	0 mmHg to 300 mmHg
Pressure Resolution	1 mmHg
Maximum Mean Error	±5 mmHg
Maximum Standard Deviation	8 mmHg
Maximum Measuring Period	Adult/ Pediatric: 120 s Neonate: 90 s
Typical Measuring Period	iCUFS measurement: 20 s to 35 s iFAST measurement: 15 s
Dual Independent Channel Overpressure Protection	Adult: (297±3) mmHg Pediatric: (245±3) mmHg Neonate: (147±3) mmHg

CNBP

Measuring Range (Adult)	SYS: 25 mmHg to 290 mmHg DIA: 10 mmHg to 250 mmHg
Measuring Range (Pediatric)	SYS: 25 mmHg to 240 mmHg DIA: 10 mmHg to 200 mmHg
Alarm Type	SYS, DIA
Pressure Resolution	1 mmHg
Maximum Mean Error	±5 mmHg
Maximum Standard Deviation	8 mmHg

BPVI

Measuring Range	0~100%
Resolution	1%
Update Frequency	5 s

EDAN Module SpO_2

Measuring Range	0% to 100%
Resolution	1%
Data Update Period	1 s
Accuracy	Adult/Pediatric: ±2% (70% to 100% SpO_2) Undefined (0% to 69% SpO_2) Neonatal: ±3% (70% to 100% SpO_2) Undefined (0% to 69% SpO_2)

PI (Perfusion Index)

Measuring Range	0 to 20%, invalid PI value is -?-.
Resolution	1% (10% to 20%) 0.1% (1.0% to 9.9%) 0.01% (0.00% to 0.99%)

Nellcor Module SpO_2

Measuring Range	1% to 100%
Resolution	1%
Data Update Period	1 s
Accuracy	
DS-100A, OXI-A/N (Adult), D-YS (Adult and Pediatric), OXI-P/I (Pediatric)	±3% (70% to 100% SpO_2)
MAX-A, MAX-AL, MAX-N, MAX-P, MAX-I, MAX-FAST (Adult and Pediatric)	±2% (70%～100% SpO_2)
MAX-A, MAX-AL, MAX-N, MAX-P, MAX-I, MAX-FAST (Adult and Pediatric)	±3% (60%～80% SpO_2)

PR

PR (SpO_2)

Measuring range	EDAN: 25 bpm to 300 bpm
	Nellcor: 20 bpm to 300 bpm
Accuracy	EDAN: ±2 bpm
	Nellcor: ±3 bpm (20 bpm to 250 bpm)
Resolution	EDAN: 1 bpm
	Nellcor: 1 bpm

PR (NIBP)

Measuring range	EDAN: 40 bpm to 240 bpm
Accuracy	EDAN: ±3 bpm or 3.5%, whichever is greater
Resolution	EDAN: 1 bpm

PR (IBP)

Measuring range	EDAN: 20 bpm to 300 bpm
Accuracy	EDAN: 30 bpm to 300 bpm: ±2 bpm or ±2%, whichever is greater; 20 bpm to 29 bpm: undefined
Resolution	EDAN: 1 bpm

TEMP

EDAN Module

Channel	2
Sensor Type	YSI-10K and YSI-2.252K
Technique	Thermal resistance
Measure Parameter	T1, T2, TD (the absolute value of T2 minus T1)
Position	Skin, oral cavity, rectum
Unit	°C, °F
Measuring Range	0°C to 50°C (32°F to 122°F)
Resolution	0.1°C (0.1°F)
Accuracy	±0.3°C (± 0.1 °C exclude sensor error)
Transient Response Time	≤30 s

Covidien Genius 3 Tympanic Thermometer

Measuring Range	33.0 °C to 42.0 °C (91.4 °F to 107.6 °F)
Accuracy	Ambient Temperature: 16 °C to 33 °C, when measuring 33 °C to 42 °C, the accuracy is ±0.3 °C.
Resolution	0.1 °C/0.1 °F
Measurement Response Time	< 2 s
Storage Temperature	-25 °C to 55 °C (-13 °F to 131 °F)
Measuring Mode	Adjusted Mode

IBP

Channel	2
Technique	Direct invasive measurement
Measuring Range	
ART, Ao, UAP, BAP, FAP, LV, P1-P4	(-50 mmHg to +400) mmHg
PA	(-6 mmHg to +120) mmHg
CVP, ICP, LAP, RAP, UVP	(-10 mmHg to +40) mmHg
Resolution	1 mmHg
Accuracy (not including sensor)	±2% or ±1 mmHg, whichever is greater ICP:0 mmHg to 40 mmHg: ±2 % or ±1 mmHg, whichever is greater; -10 mmHg to -1 mmHg: undefined
Unit	kPa, mmHg, cmH_2O

EDAN G2 Sidestream Module CO_2

Intended patient	Adult, pediatric, neonatal
Measure Parameters	$EtCO_2$, $FiCO_2$, AwRR
Unit	mmHg, %, kPa
Measuring Range	$EtCO_2$: 0 mmHg to 150 mmHg (0% to 20%)
	$FiCO_2$: 0 mmHg to 50 mmHg
	AwRR: 0 rpm to 150 rpm
Resolution	$EtCO_2$: 1 mmHg
	$FiCO_2$: 1 mmHg
	AwRR: 1 rpm
$EtCO_2$ Accuracy	
Typical conditions: Ambient temperature: (25± 3) °C Barometric pressure: (760± 10) mmHg Balance gas: N_2 Sample gas flowrate: 100 ml/min	±2 mmHg, 0 to 40 mmHg ±5% of reading, 41 to 70 mmHg ±8% of reading, 71 to 100 mmHg ±10% of reading, 101 to 150 mmHg
All conditions	±12% of reading or ±4 mmHg, whichever is greater
AwRR Accuracy	±1 rpm
Sample Gas Flowrate	50 ml/min, 70 ml/min or 100 ml/min (optional), accuracy: ±15 ml/min
Warm-up Time	Display reading within 20 s; reach to the designed accuracy within 2 minutes.
Response Time	< 4 s (with 2 m gas sampling tube, sample gas flowrate: 100 ml/min&70 ml/min)
	< 5.5 s (with 2 m gas sampling tube, sample gas flowrate: 50 ml/min)
Barometric Pressure Compensation	Automatic (The change of barometric pressure will not add additional errors to the measurement values.)
Zero Calibration	Support
Calibration	Support
Apnea Alarm Delay	10 s, 15 s, 20 s(Default), 25 s, 30 s, 35 s, 40 s

Respironics Sidestream and Mainstream Module CO_2

Applicable Patient Type	Adult, pediatric and neonatal
Method	Infra-red Absorption Technique
Measure Parameters	$EtCO_2$, $FiCO_2$, AwRR
Unit	mmHg, %, kPa
Measuring Range	$EtCO_2$: 0 mmHg to 150 mmHg
	$FiCO_2$: 3 mmHg to 50 mmHg
	AwRR: 2 rpm to 150 rpm(Sidestream) 0 rpm to 150 rpm (Mainstream)
Resolution	$EtCO_2$ 1 mmHg
	$FiCO_2$ 1 mmHg
	AwRR 1 rpm
Accuracy	
$EtCO_2$	±2 mmHg, 0 mmHg to 40 mmHg
	±5% of reading, 41 mmHg to 70 mmHg

	±8% of reading, 71 mmHg to 100 mmHg
	±10% of reading, 101 mmHg to 150 mmHg
	±12% of reading, RR is over 80 rpm (Sidestream)
	There will be no degradation in performance due to respiration rate. (mainstream)
AwRR	±1 rpm
Apnea Alarm Delay	10 s, 15 s, 20 s(Default), 25 s, 30 s, 35 s, 40 s
Zero Calibration	Support
Sample Gas Flow Rate (Sidestream)	(50 ±10) ml /min
Barometric Pressure Compensation	User setup
CO_2 Rise Time/Response Time (Mainstream)	< 60 ms
Sensor Response Time (Sidestream)	<3 seconds, includes transport time and rise time

Masimo Sidestream Module CO_2

Ambient CO_2	≤ 800 ppm (0.08 vol%)
Sampling Flow Rate	(50 ± 10) sml/min
Respiration Rate	0 to 150 ± 1 breaths/min.
Calibration	No span calibration is required.
Warm-up Time	< 10 seconds
CO_2 Rise Time at 50sml/min Sample Flow	≤ 200 ms
NomoLine ISA CO_2 System Response Time	< 3 seconds
Apnea Alarm Delay	15 s, 20 s (Default), 25 s, 30 s, 35 s, 40 s
AwRR Range	0 rpm to 150 rpm
AwRR Accuracy	± 1 rpm
CO_2 Accuracy	
Standard Conditions	±(0.2 vol% + 2% of reading), (0 to 15) vol% Unspecified, (15 to 25) vol%
All Conditions	±(0.3 kPa + 4% of reading)

Masimo Mainstream Module CO_2

Respiration Rate	0 to 150 ± 1 breaths/min.
Calibration	No span calibration required for the IR bench.
Warm-up Time	< 10 seconds
Rise Time (@ 10 l/min)	≤ 90 ms
Total System Response Time	Total system response time
Apnea Alarm Delay	15 s, 20 s (Default), 25 s, 30 s, 35 s, 40 s
AwRR Range	0 rpm to 150 rpm
AwRR Accuracy	± 1 rpm
CO_2 Accuracy	
Standard Conditions	±(0.2 vol% + 2% of reading), (0 to 15) vol% Unspecified, (15 to 25) vol%
All Conditions	±(0.3 kPa + 4% of reading)

C.O.

Technique	Thermodilution Technique
Measure Parameters	C.O., TB, TI
Measuring Range	C.O.: 0.1 L/min to 20 L/min
	TB: 23°C to 43°C (73.4ºF to 109.4ºF)
	TI: -1°C to 27°C (30.2ºF to 80.6ºF)
Resolution	C.O.: 0.1 L/min
	TB, TI: 0.1°C (+0.1ºF)
Accuracy	C.O.: ±5% or ±0.2 L/min, whichever is greater
	TB: ±0.1°C (±0.18 ºF) (not including sensor)
	TI : ±0.1°C (±0.18 ºF) (not including sensor)

Safety Specifications

Compliant with Standards	IEC 60601-1: 2005+A1 :2012; IEC 60601-1-2: 2014; EN 60601-1: 2006+A1 :2013; EN 60601-1-2: 2015; IEC 80601-2-49: 2018
Anti-electroshock Type	With AC power supply: Class I equipment With DC power supply: Class II equipment
Anti-electroshock Degree	CF
Ingress Protection	IP22

Environmental Specifications

Temperature	Working : +0°C to +40°C (32ºF ~ 104ºF)
	Transport and storage: -30°C to +70°C (-22ºF ~ 158ºF)
Humidity	Working: 15%RH to 95%RH (non-condensing)
	Transport and storage: 10%RH to 95%RH (non-condensing)
Altitude	Working: 57 kPa to 107.4 kPa
	Transport and storage: 16 kPa to 107.4 kPa

EDAN